

08/12/2023

ADDENDUM - 01

Tender no.2023-24/06

1. GeM Bid No. GEM/2023/B/4259806: Thermo Gravimetric Analyzer (TGA)

Sr. No	Tender Specifications	Amended as
1	Balance Measurement Range: Upto 1 g or more	Balance measurement Range: Upto 500 mg or more
2	A suitable inbuilt chiller with CE certification should be provided	A suitable inbuilt chiller/ water circulating system should be provided


Principal Director (T)