

18/10/2023

ADDENDUM - 02

Tender no.2023-24/03

GeM Bid No. GEM/2023/B/4019459: Computerized Universal Testing Machine - 100 kN

Sr. No	Description	Tender Specifications	Amended as
10	Additional requirement		Frame Stiffness – 250 KN/mm or better
			Cross Head Position Accuracy: 0.05 % or better
10.1		10.1 Grip head arrangements for the tensile Specimen of metal pipes with throat area available with Shank length of different diameters: - Up to 25 mm	10.1 Grip head arrangements for the tensile Specimen of metal pipes with throat area available with Shank length of different diameters: - Up to 25 mm specimen thickness




Principal Director (T)