

CORIGENDUM

Tender no. CIPET_HO/Pur/Ten.01/2020-21 dt 08/04/2020

1. Equipment Name: **Climatic Chamber**

S. No.	Description	Tendered Specification	Amended Specification
1	Humidity Range (%)	0-99	10% - 95%

2. Equipment Name: **IZOD / Charpy Impact Testing Machine**

S. No.	Description	Tendered Specification	Amended Specification
1	Optional Accessories	(i)Should offer facility for Zero & subzero testing upto – 600C	(i)Should offer facility for Zero & subzero testing upto – 60°C

3. Equipment Name: **Dielectric Breakdown Tester**

S. No.	Description	Tendered Specification	Amended Specification
1	Short Circuit Current	10, 15, 20, 25, 30, 35, 40, & 50 A	10, 15, 20, 25, 30, 35, 40, & 50 mA

4. Equipment Name: **DMA and TGA**

S. No.	Description	Tendered Specification	Amended Specification
1	Frequency Range	0.001-300Hz	0.001 - 200Hz

5. Equipment Name: **Gas Chromatography (GC)**

S. No.	Description	Tendered Specification	Amended Specification
1.	Oven	Cool down time from 450 deg. C to 50 deg. C in less than 5 minutes.	Cool down time from 450 deg. C to 50 deg. C in less than 5 minutes.
2.	Thermal Conductivity Detector	Operating temperature: 100 ^o C to 450 ^o C in steps of 10 ^o C.	Operating temperature: 100 ^o C to 400 ^o C in steps of 10 ^o C.
3.	Head Space Sampler	Sample Tray have Capacity of more than 40 Samples.	Sample Tray have Capacity of 12 Vials or more.