

CORRIGENDUM No. 4

Ref : Tender no. CIPET_HO/Pur/Ten.01/2020-21 dt 08/04/2020

SPECIFICATIONS - TESTING EQUIPMENTS

1. Equipment Name: **Izod & Charpy Impact Tester**

S. No.	Description	Tendered Specification	Amended Specification
1	Pendulum	should offer facility for zero & subzero testing upto -600C	should offer facility for zero & subzero testing upto -60°C

2. Equipment Name: **Universal Testing Machine**

Description	Tender Specification	Amended specification
Extensometer	Advanced Video Camera Extensometer - Non Contact Strain gauge	Advanced Video Camera Extensometer - Non Contact Strain gauge with Gauge Length: 10 mm - 100 mm; 200 mm field of view
Environmental Chamber	Environmental Conditioning Chamber temp. range : - 100° C to 300° C	Environmental Conditioning Chamber temp. range : - 100° C to 300° C. Chamber size should accommodate Grips and Fixtures mentioned in S No 7 and test standard requirements mentioned in S No 8.