

CORRIGENDUM No. 3

Tender no. CIPET_HO/Pur/Ten.01/2020-21 dt 08/04/2020

1. Equipment Name: **Volume and Surface Resistance tester**

S. No.	Description	Tendered Specification	Amended Specification
1	Test Voltage	10 to 1000 V AC	10 to 1000 DC
2	Type of Specimen		Solid specimens as per ASTM D 257

2. Equipment Name: **Climatic Chamber**

Tender Specification	Amended specification
Point No. 4: at -10 ° C (min) - 5 Point No.5: at 70 ° C (Min) - 1 Point No.6: at 150 ° C (min) - 5 Point No.7: Mean warm-up rate acc. to factory standard (°C/min.) -40 °C to 180 °C - 4,6 Point No.8 Mean cooling rate acc. factory standard (°C/min.) 180 °C to -40 °C - 4,1.	Recovery Time after door opening at various temperatures (min.) at -10 °C (min.) 5 at 70 °C (min.) 1 at 150 °C (min.) 5 Mean warm-up rate acc. to factory standard (°C/min.) -40 °C to 180 °C : 4 to 6 (°C/min.) Mean cooling rate acc. factory standard (°C/min.) 180 °C to -40 °C : 4 to 1 (°C/min.)

3. Equipment Name: **Universal Testing Machine**

Tender Specification	Amended specification
Horizontal day light : 400 mm	Clearance between the columns is 400 mm (Minimum)

4. Equipment Name: **Deep freezer**

Tender Specification	Amended specification
1) Model: Vertical / Horizontal 2) Volume: 500lt/1000 ltr	1) Model: Horizontal 2) Volume: 1000 litres

5. Equipment Name: **Film Cutting & Sealing Machine**

S. No.	Description	Tendered Specification	Amended Specification
1	Max Sealing Cutting Size	Max.600mm	Max. 700mm

6. Equipment Name: **Scrap Grinder – 30-50 kg/hr**

S. No.	Description	Tendered Requirement	Amended Requirement
1	Scrap Grinder – 30-50 kg/hr	CIPET Dehradun	CIPET Agartala

7. Equipment Name: **Scrap Grinder – 100 kg/hr**

S. No.	Description	Tendered Requirement	Amended Requirement
1	Scrap Grinder – 100 kg/hr	CIPET Agartala	CIPET Dehradun