

CENTRAL INSTITUTE OF PLASTICS ENGINEERING & TECHNOLOGY

Department of Chemicals & Petrochemicals,
Ministry of Chemicals & Fertilizers (Govt. of India)
B-25, C.N.I. Complex, Patia, Bhubaneswar - 751 024 (Odisha)
Phone No. 91-674-2743462, 2743767 Fax No. 91-674-2743863
E-mail: cipetbbsr@gmail.com

Tender Document
For
Dichloromethane bath Apparatus as per IS 13592:2013

Tender form Issued to:

Last date & Time for submission of Tender: 03.02.2020,

FY-2019-20

TENDER NOTICE

Date: 15.01.2020

Subject: Dichloromethane bath Apparatus as per IS 13592:2013

CIPET - IPT , Bhubaneswar invites sealed quotations from registered and authorized supplier for “**Dichloromethane bath Apparatus as per IS 13592:2013**” to be supplied at B-25, CNI Complex, Patia, Bhubaneswar. The sealed bidding documents duly filled in as per the instructions of the Tender Document addressed to the Principal Director & Head IPT, CIPET Bhubaneswar and reach latest by **17.00 Hrs. on 03.02.2020**. Tender Documents and detail may be obtained from CIPET-IPT Bhubaneswar office during working day between 10.00 AM to 4.00 PM or downloaded from the Website (www.cipet.gov.in).

Tender Reference	CIPET/IPT/BBS/DM/01 Dt: 15.01.2020
Last Date and Time for submission	03.02.2020., 17.00Hrs
Time and date of opening of bid	04.02.2020, 3.00 PM
Address for Communication	As mentioned below
Contact TelephoneNumbers	0674-2743 462/2744 607

The bid should be in a big envelope duly subscribed with tender reference no. and tender notice detail as appended here under: -

**Dichloromethane bath Apparatus as per IS: 13592:2013 TENDER NO. CIPET/IPT/BBS/DM
Dt: 15.01.2020**

LAST DATE FOR SUBMISSION: 03.02.2020

To

The Principal Director & Head, IPT
Central Institute of Plastics Engineering & Technology
B-25, CNI Complex, Patia,
Bhubaneswar - 751 024. Odisha, India.

From: M/s
Contact No:
E-mail ID:

Technical Specification:

Dichloromethane bath Apparatus as per IS 13592:2013

<u>S.No</u>	<u>Specification Details</u>
--------------------	-------------------------------------

1. Machine design and construction as per IS: 13592; 2013
2. Grating so fitted that it is approximately 1 cm above the bottom of the container.
3. Lid to limit the evaporation of the dichloromethane.
4. Thermostat, to maintain the temperature of the dichloromethane at $15 \pm 0.5^{\circ}\text{C}$.
5. Cooling device to cool the dichloromethane to the specified temperature.
6. Stirrer to homogenize the temperature of the bath.

General Terms & Conditions:

1. The above quantity will be supplied within 20 days from the date of issue of Purchase Order.
2. Payment terms: 100% payment after successful installation and satisfactory performance as per required standard.
3. Order will be placed based on the L1 party from the quoted price.
4. Party should have valid GST & PAN and copy of the same should be submitted along with the tender.
5. The competent authority has the right to reject any or all quotations without assigning any reason thereto.

Dichloromethane bath Apparatus as per IS 13592:2013

1. THE COMPANY

a) Name _____

b) Regd. Address _____

c) Address of Local Office _____

d) Contact Person's

i) Name & Design. _____

ii) Tel No. Landline _____ Mobile _____

iii) Email ID _____

2. Type of Firm : Private Ltd./Public Ltd./PSU
(Please tick and enclose copy of Memorandum/Articles of Association/Certificate of Incorporation)

3. PAN No.: _____
(Please enclose attested photocopy)

4. GST No.: _____
(Please enclose attested photocopy)

6. Experience of similar work in the field during the last three assignments

Signatures of authorized signatory

Name _____

Designation _____

Seal:

Dichloromethane bath Apparatus as per IS 13592:2013

Sl. No.	Details of Scope of work	Qty.	Amount/Rs. Per unit
01.	Dichloromethane bath Apparatus as per IS:13592:2013	01 No.	Rs. _____
02	Other Accessories		
02	GST		Rs. _____
03	Any other charges (if any) Please specify:		Rs. _____
Total			Rs. _____
In word (Rupees _____ Only)			

Signatures of authorized person

Name _____

Designation _____

Seal