

CORRIGENDUM - II

With reference to the e-Tender Notice No.06/2021-22 the following changes are made as under:-

1. For **Supply of Inverted Fluorescent Microscope with Automated Z Focus for Advance Fluorescence Imaging** at CIPET:SARP-LARPM, Bhubaneswar:-

Sl. No	Existing Specification	Revised specification
1.	Motorized inverted Microscope stand with apochromatically corrected fluorescence beam path with transmitted light LED illumination for bright field, phase contrast, fluorescence, DIC and upgradable to dark field option for Bio-Medicine Applications.	Motorized inverted Microscope stand with apochromatically corrected fluorescence beam path/equivalent technology with transmitted light LED illumination for brightfield, phase contrast, fluorescence, DIC and upgradable to darkfield option for Bio-Medicine Applications.
	The microscope should be provided with left and right side for attachment of two cameras.	The microscope should be provided with two ports for attachment of two cameras.
11	FOV for camera should be minimum 23 mm.	FOV for camera should be minimum 19 mm

2. The last for receiving of e-Tender for both of the Equipments is hereby extended upto **18.09.2021 till 03.00 P.M**

All other terms & conditions of the e-Tender remains unchanged.

Asst. Admin. Officer