

11/03/2022

ADDENDUM 01

Tender no.2021-22/06

Equipment: 2021-22/06/03 – Capillary Rheometer

| Sl. No. | Item | Specification | Revised specification |
|---------|--|-----------------|--|
| 6 | Maximum Force Range | 25 KN or better | 50 KN or better |
| 16 | Rheological corrections & measurements | | Provision for upgrading the system as per ASTM D 5930 for Thermal Conductivity Electronics, Probe, etc |

***Note: Against sl. No. 16 (Provision for upgrading the system as per ASTM D 5930 for Thermal Conductivity Electronics, Probe, etc) to be added.**



Principal Director (T)

