

Corrigendum

GC-MS with Pyrolyzer and Headspace

Point No.	Parameter	Tender Specification	Corrigendum
<u>General Technical</u>	Data acquisition /operating system	System should be capable of supporting three inlets and three detector ports simultaneously	System should be capable of supporting inlets for Pyrolyzer, PTV, and Headspace; and; three detector ports simultaneously.
<u>GC Detector Specification</u>	Carrier gas head pressure setting	Should be more than 950 KPa	650 KPa or better for PTV; and 950 KPa or better for split/splitless injection ports
	Turbo molecular pump (TMP)	300 L/s or better capacity	Minimum 200 L/s or higher


 Principal Director (T)