

23/01/2024

ADDENDUM - 02

Tender no.2023-24/06

1. GeM Bid No. GEM/2023/B/4382478: ICP - MS

Sr. No	Tender Specifications	Amended as
7	Mass range: The mass range should be from 5-280 amu or broader.	Mass range: The mass range should be from 5-260 amu or broader.
11	Microwave digestion system for sample preparation: The system should have 14 or more vessel rotor	The system should have 10 or more vessel rotor


Principal Director (T)